

**10A NCAC 42D .0246 MICROBIOLOGY FACILITIES**

Laboratory facilities shall be clean, temperature and humidity controlled, and have sufficient lighting at bench tops to perform the required procedures. The laboratory shall have provisions for disposal of microbiological waste.

*History Note: Authority G.S. 130A-315;  
Eff. December 1, 1991;  
Pursuant to G.S. 150B-21.3A, rule is necessary without substantive public interest Eff. December 23, 2017.*