

10A NCAC 42D .0258 RADIOCHEMISTRY SAMPLE COLLECTION/HANDLING/PRESERVATION

- (a) Minimum requirements of sample handling for radiochemical analyses including sample container, preservation, and major instrumentation shall meet the criteria in Table VI-I in the EPA "Manual for the Certification of Laboratories Analyzing Drinking Water", which is hereby incorporated by reference including any subsequent amendments and editions. A copy is available for inspection at the Division of Public Health, 1915 Mail Service Center, Raleigh, North Carolina 27699-1915. Nonprofit organizations or government agencies may obtain a copy by contacting the EPA Drinking Water Hotline at 800-426-4791. Other organizations may obtain a copy from the National Technical Information Service at 800-336-4700 for thirty five dollars (\$35.00).
- (b) Samples shall be collected in containers provided by the laboratory.
- (c) If composited samples are to be analyzed, the compositing shall be performed in the laboratory.
- (d) Sample report form. The sample report form shall contain the location; date and time of collection; collector's name; preservative added; and any other special remarks concerning the sample. Sample report forms shall be approved by the N.C. Public Water Supply Section. Indelible ink shall be used to complete the form.

History Note: Authority G.S. 130A-315;
Eff. December 1, 1991;
Pursuant to G.S. 150B-21.3A, rule is necessary without substantive public interest Eff. December 23, 2017.