

21 NCAC 22I .0110 STANDARDS FOR AUDIOMETERS

All audiometers used for fitting and selling hearing aids shall be in calibration according to the standards contained in the ANSI S3.6-1969 specifications for audiometers which are herein adopted by reference, in accordance with G.S. 150B-14 (b).

History Note: *Authority G.S. 93D-3(c); 150B-14;*
 Eff. April 23, 1976;
 Amended Eff. May 1, 1988;
 Pursuant to G.S. 150B-21.3A, rule is necessary without substantive public interest Eff. September 6,
 2016.