

21 NCAC 46 .2706 PROHIBITIONS

(a) No person shall utilize unit-dose transport containers for radioactive dosages without an effective mechanism to avoid contamination of the transport container with blood or other biohazardous substances.

(b) No person shall re-use a unit-dose transport container that has been contaminated with blood or other biohazardous substances. Any unit-dose transport container that is returned with the tamper-evident seal broken and the unit-dose syringe included must be considered to be contaminated.

History Note: *Authority G.S. 90-85.6; 90-85.34;*
 Eff. January 1, 2005;
 Pursuant to G.S. 150B-21.3A, rule is necessary without substantive public interest Eff. October 3,
 2017.