

**10A NCAC 13B .4902    RECORDS**

- (a) All requests for laboratory services shall be documented.
- (b) All reports of laboratory services performed, including autopsy, shall be placed in the patient's medical record.
- (c) Records of proficiency testing appropriate to the scope of services offered shall be available to the Division for review.
- (d) Records of equipment calibration and quality controls as recommended by the manufacturer shall be maintained and be available to the Division for review.

*History Note:*     *Authority G.S. 131E-79;*  
                      *Eff. January 1, 1996;*  
                      *Pursuant to G.S. 150B-21.3A, rule is necessary without substantive public interest Eff. July 22,*  
                      *2017.*