

10A NCAC 15 .0110 PROHIBITED USES

- (a) Hand-held fluoroscopic screens shall not be used.
- (b) Shoe-fitting fluoroscopic devices shall not be used.
- (c) Effective February 1, 1981, plastic pointed position indicating devices on intraoral dental systems shall not be used.
- (d) Effective February 1, 1983, mechanical timers on intraoral dental machines shall not be used.
- (e) Dental fluoroscopy without image intensification shall not be used.
- (f) Non-intensified photofluorographic equipment shall not be used.

History Note: *Authority G.S. 104E-7;*
 Eff. February 1, 1980;
 Amended Eff. June 1, 1989;
 Transferred and Recodified from 10 NCAC 3G .2211 Eff. January 4, 1990;
 Transferred and Recodified from 15A NCAC 11 .0110 Eff. February 1, 2015.