

**10A NCAC 15 .0359 MEASUREMENTS/DOSAGES OF UNSEALED RADIOACTIVE MATERIAL FOR MEDICAL USE**

(a) Licensees shall comply with the provisions of 10 CFR 35.63, except that dosage determination shall be made by direct measurement for all unsealed photon-emitting radioactive drugs prior to administration to any person. Licensees shall ensure that instruments used to measure dosages under this Rule are calibrated in accordance with the provisions of 10 CFR 35.60.

(b) Notwithstanding Rule .0117 of this Chapter, the regulations cited in this Rule from 10 CFR Chapter I (2015) are hereby incorporated by reference, excluding subsequent amendments and editions. Copies of these regulations are available free of charge at [http://www.ecfr.gov/cgi-bin/text-idx?SID=2beece594411a03e50b2468ae31f89b&pid=20160101&tpl=/ecfrbrowse/Title10/10tab\\_02.tpl](http://www.ecfr.gov/cgi-bin/text-idx?SID=2beece594411a03e50b2468ae31f89b&pid=20160101&tpl=/ecfrbrowse/Title10/10tab_02.tpl).

*History Note:* Authority G.S. 104E-7; 104E-10(b); 104E-12;  
Eff. April 1, 1999;  
Amended Eff. November 1, 2007;  
Transferred and Recodified from 15A NCAC 11 .0359 Eff. February 1, 2015;  
Amended Eff. March 1, 2017.